Europäisches Patentamt

(19) **European Patent Office**

Office européen des brevets

(11) Publication number : **0 506 302 B1**

(12) # EUROPEAN PATENT SPECIFICATION

(45) Date of publication of patent specification :
08.11.95 Bulletin 95/45

(51) Int. Cl.⁶ : **G01R 33/563, G06T 11/00**

(21) Application number : 92302390.7

(22) Date of filing : 19.03.92

(54) Projection methods for producing two-dimensional images from three-dimensional data.

(30) Priority : 25.03.91 US 674476

(43) Date of publication of application :
30.09.92 Bulletin 92/40

(45) Publication of the grant of the patent :
08.11.95 Bulletin 95/45

(84) Designated Contracting States :
DE FR GB NL

(56) References cited :
EP-A- 0 366 387
US-A- 4 984 160
MAGNETIC RESONANCE IN MEDICINE. vol.
14, no. 2, May 1990, DULUTH, USA pages 222 -
229; G. LAUB: 'Displays for MR Angiography'
9TH INTERNATIONAL CONFERENCE ON PAT-
TERN RECOGNITION vol. 2, November 1988,
ROME, ITALY pages 689 - 692; T. JIANG ET
AL.: 'Automated Threshold Detection Using a
Pyramid Structure'

(56) References cited :
MAGNETIC RESONANCE IN MEDICINE. vol.
11, no. 1, July 1989, DULUTH, USA pages 35 -
46; C.L. DUMOULIN ET AL.: 'Three-Dimen-
sional Time-of-Flight Magnetic Resonance
Angiography Using Spin Saturation'
MAGNETIC RESONANCE IN MEDICINE. vol. 9,
no. 1, January 1989, DULUTH, USA pages 139 -
149; C.L. DUMOULIN ET AL.: 'Three-Dimen-
sional Phase Contrast Angiography'

(73) Proprietor : GENERAL ELECTRIC COMPANY
1 River Road
Schenectady, NY 12345 (US)

(72) Inventor : Souza, Steven Peter
136 Lindley Place
Williamstown, Massachusetts 01267 (US)
Inventor : Adams, William John
6 Valdepenas Lane
Clifton Park, New York 12065 (US)
Inventor : Dumoulin, Charles Lucian
36 Terrace Court
Ballston Lake, New York 12019 (US)

(74) Representative : Greenwood, John David et al
London Patent Operation,
General Electric Technical Services Company,
Inc.,
Essex House,
12/13 Essex Street
London WC2R 3AA (GB)

EP 0 506 302 B1

Description

BACKGROUND OF THE INVENTION

1. Field of the Invention

This invention relates generally to Magnetic Resonance Angiography signal processing and more specifically to a method of reducing three-dimensional Magnetic Resonance Angiography or stationary tissue data to a two-dimensional image using statistical manipulation, while reducing noise and/or increasing contrast.

2. Description of Related Art

Known techniques for magnetic resonance angiography include methods that, either by volume or sequential-slice acquisition, produce data sets with spatial resolution in three dimensions. These techniques may utilize either time-of-flight effects or motion-induced phase shifts to produce flow contrast. Since there is at this time no practical real-time three-dimensional display technology (e.g., holographic video) available, this data is reduced to two spatial dimensions for convenient viewing by the radiologist. This has generally been accomplished by generating a series of projection images at sequential view angles. These images are then repeatedly displayed in rapid sequence on a display screen to give the viewer a three-dimensional impression of the vascular anatomy. The projection images are created by ray tracing through the volume of data, using some predetermined method for calculating the projected pixel value from each ray. In the case of phase contrast data, it is desirable to preserve the quantitative nature of the data insofar as is possible.

The most obvious projection method is to take the mean of all volume pixel ("voxel") values along each cast ray. This unfortunately yields images with low vessel contrast and poor detection of fine features and vessels, especially for time-of-flight data that always contain signals from unsuppressed stationary tissue. This is because vessels are sparse within the data volume and the total noise and/or background signal along a ray can easily be comparable to or much larger than the total flow signal along that ray. To provide good vessel contrast, the method of choice has become projection of the largest voxel value along each ray. (See Dumoulin, C.L., Souza, S.P., Walker, M.F., and Wagle, W., "Three-dimensional phase contrast angiography," *Magn. Reson. Med.*, (1989), and Armatur, S.C., Masaryk, T. J., Modic, M.T.; et al., "3DFT time-of-flight magnetic resonance angiography," *Dynamic Cardiovascular Imaging*, 2. 170, 1989). This method is generally known as Maximum Intensity Projection (MIP). The MIP method uses the maximum voxel value along a projection ray

and discards the remaining voxel values along that ray. Since along each ray which actually intercepts one or more vessels the only voxel which contributes to the projected pixel is within a vessel, the contrast is very good. Unfortunately MIP has several drawbacks. Since the MIP method retains only one voxel along each ray, it does not make efficient use of the acquired data. This is especially true where a vessel extends over many voxels and averaging over the vessel might be employed to improve the vessel signal-to-noise ratio. The MIP method tends to lose contrast for vessels with strong signal near the walls but weak signal near the axis of the vessel. This situation occurs in phase contrast MR angiography when the velocity at the center of the vessel exceeds the maximum for which the applied flow encoding produces monotonically increasing signal. The MIP method also may obscure real pathological features. For example, this will occur when a stenotic lesion distorts a vessel wall and the lesion appears along the line-of sight rather than displayed in profile. There is a need for a projection method that retains the maximum amount of information possible when transforming the acquired data to projection data while introducing no artifactual features and discarding few actual features.

OBJECTS OF INVENTION

It is an object of the present invention to construct a high-contrast low-noise two-dimensional image from three-dimensional magnetic resonance (MR) fluid flow data from a volume of a subject.

It is another object of the present invention to alter the contrast between structures of a subject imaged to make differing structures more visible.

It is another object of the present invention to adapt the MR data to reduce differing types of noise inherent in the MR data.

It is another object of the present invention to construct projection images from three-dimensional MR stationary tissue data, thereby producing synthetic images similar to conventional x-ray images but with alterable contrast between structures.

SUMMARY OF THE INVENTION

In this invention, three-dimensional Magnetic Resonance fluid flow or stationary tissue measurements of a volume of a subject are processed into high-contrast low-noise two-dimensional images. The first step is to acquire by known means a three-dimensional data set having a number of entries, each entry corresponding to a Magnetic Resonance fluid flow or stationary tissue volume element, or volume pixel (voxel) value of said subject. The value of each pixel in the final two-dimensional image is a calculated statistic of the array of voxel values intercept-

ed by an imaginary ray cast through the three-dimensional data along the line of sight to that particular pixel at some view angle. An order of the statistic is chosen which depends on the general shape of a histogram of the data that is to be imaged. A projection ray is passed at a predetermined angle through the data volume and the three-dimensional data set entries corresponding to all intersected voxels are chosen. The average and standard deviations of this set of entries are taken.

The projection average is subtracted from each of the entries and the result is then raised to a power being equal to that of a statistical order, to obtain the weighted voxel differences.

The sum of the weighted voxel differences is divided by the number of entries, then divided by a normalization constant to result in a projection value. This procedure is repeated for all parallel rays covering the three-dimensional data set and intersecting all pixels in the projection image, and the data set is then displayed as a two-dimensional fluid flow image.

BRIEF DESCRIPTION OF THE DRAWINGS

The features of the invention believed to be novel are set forth with particularity in the appended claims. The invention itself, however, both as to organization and method of operation, together with further objects and advantages thereof, may best be understood by reference to the following description taken in conjunction with the accompanying drawings(s) in which:

Fig. 1 is a block diagram showing the apparatus of the present invention that is used for obtaining the three-dimensional MR data, and constructing a two-dimensional projection image.

Fig. 2 is a graphical representation of the geometry employed by the present invention, showing a three-dimensional data set volume, a projection ray, and a weighted projection value pertaining to the projection ray.

Fig. 3 shows a volume of the subject to be imaged that corresponds to the data set volume shown in Fig. 2.

Fig. 4 is a histogram of the three-dimensional MR data obtained by a typical phase contrast fluid flow image shown in normal amplification and 50 times amplification.

Fig. 5 is a histogram of a typical "time-of-flight" fluid flow image.

Fig. 6 is a flow chart illustrating use of the present invention.

DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENT

U.S. Patents No. 4,431,968 Edelstein et al. Issued Feb. 14, 1984, and 4,796,635 Dumoulin issued Jan. 10, 1989 and assigned to the instant assignee disclose an apparatus such as that shown in Fig. 1 that is used to create three-dimensional magnetic resonance images of fluid flow or stationary tissue in a subject. Magnet 428 immerses the subject to be imaged in a static magnetic field. Computer 401 sends gradient power pulses via interface 405 to gradient power supplies 406, 408, and 410. Gradient power supply 406 energizes gradient coil 416. The time-changing current passing through gradient coil 416 induces a time-changing magnetic field having a gradient in the x direction. Similarly, current passing through coils 418 and 420 produce magnetic fields having gradients in the y and z directions respectively.

Computer 401 sends RF excitation pulse information to transmitter 402 via interface 405. Transmitter 402 creates an RF pulse of a predetermined strength, duration, and frequency that is amplified via RF power amplifier 412. The RF pulse is transmitted by transmitter coil 424 to the subject. The combination of the RF pulse, the static magnetic field, and the magnetic field gradients causes magnetic resonance of nuclei of the subject that can be localized.

The signals received from the subject are received by a receiver coil 426, then processed by a preamplifier 422, receiver 414, and averager 404. The signal is then sent through an interface 405 to disk storage 403. The entire set of signals is received and stored. The signals are reconstructed by computer 401 into a three-dimensional data set volume 41 as shown in Fig. 2. This three-dimensional data set volume 41 is composed of voxels, each corresponding to fluid flow approximately as it occurs in the original volume 54 of Fig. 3.

After reconstructing the three-dimensional data set 41, computer 401 represents portions of the data on two-dimensional display device 430, via interface 405. The display device may be a printer, plotter, color-coded color monitor, or gray scale monitor.

For a more detailed description of operation of the apparatus shown in Fig. 1, see U.S. Patent No. 4,431,968.

To display an image on a two-dimensional screen, a set of projection amplitudes 45 of Fig. 2 must be taken of the three-dimensional data set volume 41. Casting projection rays 48 through the entire three-dimensional data set volume 41 produces a set of projection amplitudes 45. Each projection amplitude 45 is created by processing a plurality 43 of data set entries 46, each representing a corresponding voxel volume of the subject, along projection ray 48.

The prior art would simply average all data set entries 46 along projection ray 48 to obtain the projection amplitude 45 pertaining to that projection ray 48.

Another method, the Maximum Intensity Projection (MIP) method, is applied by simply taking the value of the voxel along projection ray 48 having the

highest amplitude and using this value as the projection amplitude 45 of projection ray 48.

The present invention employs a method of statistical calculation of data set entries 46 along projection ray 48 permitting an adjustable tradeoff between increased contrast (i.e., between small vessels 51 and background signal as shown in Fig. 3), and an increased signal to noise ratio. This statistical method employs an adjustment parameter k, the order of the statistic, which is adjusted to image differing types of data.

Consider a histogram of the voxel values in a three-dimensional MR angiogram, as shown in Fig. 4 for phase contrast data and Fig. 5 for time-of-flight data, in each case for one set of rays from a three-dimensional data set. In the case of phase contrast data there are at least two components to the distribution: a large, symmetrical, nearly Gaussian peak 11 at low voxel values representing noise, and a tail 12 extending toward high voxel values representing true vessel information (possibly with a contribution from faint ghosts of the vessels). The noise in peak 11 dominates the mean of this distribution (its lowest-order statistic), with only a small contribution from voxel values 12.

The MIP method takes the maximum voxel value (V_{max}) along projection ray 48 of Fig. 2. This single voxel value V_{max} is chosen, and all other voxel values are discarded.

The present invention uses an approach that is different from either of the above methods. The present invention uses intermediate-order statistics representative of the distribution of voxel values along each projection (e.g.,48 of Fig. 2). Standard deviation (or its related measure variance), skewness, kurtosis, etc. are defined by:

$$\frac{1}{N\sigma^{k-1}} \sum^{N} (V-V_i)^k$$

where N is the number of samples, V is the mean of the distribution, k is the order of the statistic, σ is the standard deviation of the distribution and V_i is an individual voxel value intersected by projection ray 48 of Fig. 2. The resulting values are normalized to provide the same units as V_i. In addition, k=2 corresponds to standard deviation, that is a measure of the width of a distribution, k=3 corresponds to skewness (a measure of the asymmetry of a distribution), and k=4 corresponds to kurtosis (a measure of the flatness or peakedness of a distribution). Higher orders of k may be defined, as well as non-integral orders (e.g., k=2.7). The MIP method is roughly equivalent to k=infinity.

Each of these statistical measures will indeed detect vessels and other structures, but the best measure is highly data-dependent. In the distribution shown in Fig. 4, skewness is likely to be the best measure since the high voxel values 12 contribute to the asymmetry of the distribution but the low value (or noise) voxel values 11 do not. Using skewness rather than the MIP method for the data set yields projection images with improved contrast and signal-to-noise ratios for larger vessels without loss of fine detail.,.

In general, lower-order statistics, i.e., lower k values, yield higher signal-to-noise ratios, better rendition of extended structures such as medium-to-large vessels, and better sensitivity to features along the projection ray 48 of Fig. 2. They also better preserve the quantitative nature of phase contrast data. Higher-order statistics yield better contrast against noise or stationary-tissue background and better sensitivity to small (e.g., single-voxel) features occurring at the right hand side, the tail 12 of Fig. 4. As a general rule, one should use the lowest-order statistic that provides sufficient contrast and background suppression. For phase contrast data the best statistic is likely to be standard deviation or skewness. For time-of-flight data a higher-order statistic is generally better.

An example of how the present invention is to be applied is shown in Fig. 6. Data is acquired in step 18 by employing a three-dimensional flow-sensitive MR imaging technique. The apparatus of Fig. 1 is employed to collect this data and perform the computations.

The characteristics of the distribution histogram of the data, similar to Fig. 4 or 5, are constructed in step 19. The order of the statistic is determined, which reflects the shape of the data in the histogram. (e.g., if the data to be imaged adds to the skewness of the histogram, the order of the statistic should be "3".)

The set of data set entries 43, which intersect a ray 48 of Fig. 2, is determined. An average set 43 is taken in step 21 of Fig. 5. In step 22 the standard deviation of the voxels intersecting ray 48 of Fig. 2 is determined, except in the case k=2. The standard deviation obtained is raised to a power equal to the order of the statistic minus one to obtain the normalization constant at step 23. Each voxel value V_i is subtracted from the voxel average to create a distribution difference in step 24.

The distribution difference is raised to a power equal to the order of the statistic, to obtain a weighted voxel difference in step 25.

It is then determined in step 26 if the voxel value that was processed was the last voxel intersecting ray 48 of Fig. 2. If it was not, processing continues along path 27, and the next voxel along the projection ray 48 of Fig. 2 is chosen in step 28. Steps 24 and 25 are then repeated for voxels intersecting this projection ray.

If the voxel processed at step 25 was the last voxel intersecting ray 48 of Fig. 2, processing passes along path 30 to step 32. The weighted voxel differ-

ences for voxels intersecting ray 48 of Fig. 2 are summed in step 32 to obtain a projection sum. The projection sum is divided by the number of voxels intersecting ray 48 of Fig. 2, then (except in the case of k=2), divided by the normalization constant in step 33 to obtain a weighted projection value for projection rays 45 and 48 of Fig. 2. (For k=2 proper normalization at step 33 is obtained by taking the square root of the projection sum divided by the number of voxels intersecting ray 48 of Fig. 2).

In step 34 it is determined if the projection ray processed is the last projection ray. If it is not the last projection ray, the next projection ray is chosen in step 36, and processing returns to step 21 to repeat the processing for the new projection ray. Processing continues as described above until all parallel projections through threedimensional data set volume 41 (Fig. 2) are processed. The two-dimensional array of projection values is then a numerical representation of the projection image.

When all the projections have been processed, the processing continues along path 38. The projections are then transferred to an electronic two-dimensional display device (430 of Fig. 1) and are displayed in step 39 by computer 401 through the interface 405 to display device 430 of Fig. 1 device for viewing.

The methods described herein will be useful in obtaining high quality projection angiograms from three-dimensional MR flow data. These methods result in high-contrast low-noise two-dimensional images from collecting three-dimensional magnetic resonance (MR) fluid flow data from a volume of a subject.

These methods are also useful when applied to MR stationary tissue data to produce projection images similar to conventional x-ray images, but with contrast between structures that can be altered by varying the order of the statistic k. Further, these methods are not limited to data obtained by magnetic resonance methods, but may be applied to any three-dimensional image data.

While only certain preferred features of the invention have been illustrated and described herein, many modifications and changes will occur to those skilled in the art. It is, therefore, to be understood that the appended claims are intended to cover all such modifications and changes.

Claims

1. A method of displaying a two-dimensional (2D) projection image of desired structures of a subject from collected three-dimensional (3D) image data of a volume of a subject comprising the steps of :

a) acquiring a 3D data set having a plurality of entries, each entry being a value corresponding to at least one physical parameter representative of a voxel within said volume;

b) acquiring a histogram of said 3D data set and selecting a statistical order k according to the general shape of the histogram of the 3D data;

c) choosing a viewing angle from which to view said volume;

d) determining a set of intersected voxels which are intersected by a selected imaginary projection ray passing through said volume at the viewing angle and determining N intersected data set entries V_I, corresponding to the intersected voxels;

e) calculating an average V and a standard deviation σ from the N intersected data set entries V_i;

f) calculating a projection value P_k from the N intersected data set entries V_I, for k not equal to 2 according to the equation :

$$P_k = \frac{1}{N\sigma^{k-1}} \sum^N (V-V_1)^k$$

or, for k = 2, according to the equation:

$$P_2 = \frac{1}{N} \left[\sum^N (V - V_i)^2 \right]^{1/2}$$

g) repeating steps "d"-"f" for a plurality of parallel imaginary projection rays at the viewing angle covering said volume;

h) transferring the projection values P_k to a two-dimensional display device;

i) displaying a two-dimensional image created from the projection values P_k on the two dimensional display device for analysis;

j) increase the value of the statistical order k and repeat steps "d"-"i" if a greater contrast of said structures in the projection image is desired; and

k) decrease the value of the statistical order k and repeat steps "d"-"i" if a greater signal-to-noise ratio of said structures in the projection image is desired.

2. The method of displaying a 2D projection image as recited in Claim 1 wherein the 3D data set is acquired by phase contrast magnetic resonance imaging techniques.

3. The method of displaying a 2D projection image as recited in Claim 1 wherein the 3D data set is acquired by time-of-flight magnetic resonance

flow imaging techniques.

4. The method of displaying a 2D projection image as recited in Claim 1 further including the step of varying the order of the statistic for a predetermined three-dimensional data set to produce images with differing contrast in elements in said subject imaged.

5. The method of displaying a 2D projection image as recited in Claim 1 wherein the 3D data set is acquired by magnetic resonance imaging techniques used to produce images of substantially stationary tissue.

6. The method of displaying a 2D projection image as recited in Claim 1 wherein the 3D data set is acquired by tomographic imaging techniques.

7. The method of displaying a 2D projection image as recited in Claim 1 wherein the 3D data set is acquired by magnetic resonance imaging techniques.

Patentansprüche

1. Verfahren zur Darstellung eines zweidimensionalen (2D) Projektionsbildes von gewünschten Strukturen eines Subjektes aus gesammelten dreidimensionalen (3D) Bilddaten von einem Volumen eines Subjektes, enthaltend die Schritte:

a) Gewinnen eines 3D-Datensatzes mit mehreren Einträgen, wobei jeder Eintrag einen Wert hat entsprechend wenigstens einem physikalischen Parameter, der ein Voxel in dem Volumen darstellt,

b) Gewinnen eines Histogrammes des 3D-Datensatzes und Wählen einer statistischen Ordnung k gemäß der allgemeinen Form des Histogrammes der 3D-Daten,

c) Wählen eines Betrachtungswinkel, von dem das Volumen betrachtet wird,

d) Ermitteln eines Satzes von geschnittenen Voxeln, die durch einen gewählten imaginären Projektionsstrahl geschnitten sind, der bei dem Betrachtungswinkel durch das Volumen hindurchtritt, und Ermitteln N geschnittener Datensatzeinträge V_i entsprechend den geschnittenen Voxeln,

e) Berechnen eines Mittelwertes V und einer Standardabweichung σ von den N geschnittenen Datensatzeinträgen V_i,

f) Berechnen eines Projektionswertes P_k aus den N geschnittenen Datensatzeinträgen V_i für k ungleich 2 gemäß der Gleichung:

$$P_k = \frac{1}{N\sigma^{k-1}} \sum^{N} (V - V_i)^k$$

oder für k=2 gemäß der Gleichung:

$$P_2 = \frac{1}{N} \left[\sum^N (V - V_i)^2 \right]^{1/2} \frac{1}{N}.$$

g) Wiederholen der Schritte "d"-"f" für mehrere parallel imaginäre Projektionsstrahlen bei dem das Volumen überdeckenden Betrachtungswinkel,

h) Übertragen der Projektionswerte P_k zu einer zweidimensionalen Darstellungsvorrichtung,

i) Darstellen eines zweidimensionalen Bildes, das aus den Projektionswerten P_k auf der zweidimensionalen Darstellungsvorrichtung für eine Analyse hervorgerufen ist,

j) Vergrößern des Wertes der statistischen Ordnung k und Wiederholen der Schritte "d"-"i", wenn ein größerer Kontrast der Strukturen in dem Projektionsbild gewünscht wird, und

k) Verkleinern des Wertes der statistischen Ordnung k und Wiederholen der Schritte d"-"i", wenn ein größeres Signal/Rausch-Verhältnis der Strukturen in dem Projektionsbild gewünscht wird.

2. Verfahren zum Darstellen eines 2D-Projektionsbildes nach Anspruch 1, wobei der 3D-Datensatz durch Phasenkontrast-Magnetresonanz-Bildgebungstechniken gewonnen wird.

3. Verfahren zum Darstellen eines 2D-Projektionsbildes nach Anspruch 1, wobei der 3D-Datensatz durch Laufzeit-Magnetresonanzströmungs-Bildgebungstechniken gewonnen wird.

4. Verfahren zum Darstellen eines 2D-Projektionsbildes nach Anspruch 1, ferner den Schritt enthaltend, daß die Ordnung der Statistik für einen vorbestimmten dreidimensionalen Datensatz verändert wird, um Bilder mit unterschiedlichem Kontrast in Elementen des abgebildeten Subjektes zu erzeugen.

5. Verfahren zum Darstellen eines 2D-Projektionsbildes nach Anspruch 1, wobei der 3D-Datensatz durch Magnetresonanz-Bildgebungstechniken gewonnen wird, die angewendet werden, um Bilder mit im wesentlichem stationärem Gewebe zu erzeugen.

6. Verfahren zum Darstellen eines 2D-Projektions-
bildes nach Anspruch 1, wobei der 3D-Datensatz
durch tomografische Bildgebungstechniken ge-
wonnen wird.

7. Verfahren zum Darstellen eines 2D-Projektions-
bildes nach Anspruch 1, wobei der 3D-Datensatz
durch Magnetresonanz Bildgebungstechniken
gewonnen wird.

Revendications

1. Procédé pour visualiser une image en projection
bidimensionnelle (2D) de structures souhaitées
d'un objet à partir de données d'image tridimen-
sionnelles (3D) recueillies concernant un volume
d'un objet, comprenant les étapes consistant à :
    (a) acquérir un jeu de données 3D ayant une
    pluralité d'entrées, chaque entrée étant une
    valeur qui correspond à au moins un paramè-
    tre physique représentatif d'un voxel dans le-
    dit volume,
    (b) acquérir un histogramme dudit jeu de don-
    nées 3D et choisir un ordre statistique k en
    fonction de la forme générale de l'histogram-
    me des données 3D,
    (c) choisir un angle d'observation depuis le-
    quel on souhaite observer ledit volume,
    (d) déterminer un groupe de voxels intersec-
    tés qui sont coupés par un rayon imaginaire
    de projection choisi qui traverse ledit volume
    suivant ledit angle d'observation et détermi-
    ner N entrées V_i d'un groupe de données in-
    tersectées qui correspondent aux voxels in-
    tersectés,
    (e) calculer la valeur moyenne V et l'écart type
    σ des N entrées V_i du groupe de données in-
    tersectées,
    (f) calculer une valeur de projection P_k à partir
    des N entrées V_i du groupe de données inter-
    sectées,
    pour k différent de 2, d'après l'équation :

$$P_k = \frac{1}{N \, \sigma^{k-1}} \, \Sigma^N (V - V_i)^k$$

et pour k = 2, d'après l'équation :

$$P_2 = \frac{1}{N} \, \{ \Sigma^N (V - V_i)^2 \}^{1/2}$$

    (g) répéter les étapes (d) à (f) pour une plura-

lité de rayons imaginaires de projection, pa-
rallèles, faisant ledit angle d'observation et re-
couvrant ledit volume,
    (h) transférer les valeurs de projection P_k à un
    dispositif de visualisation en deux dimen-
    sions,
    (i) afficher sur le dispositif de visualisation en
    deux dimensions une image bidimensionnelle
    créée à partir des valeurs de projection P_k,
    pour analyse,
    (j) augmenter la valeur de l'ordre statistique k
    et répéter les étapes (d) à (i) si on souhaite un
    plus fort contraste desdites structures dans
    l'image de projection, et
    (k) diminuer la valeur de l'ordre statistique k et
    répéter les étapes (d) à (i) si on souhaite un
    plus fort rapport signal/bruit desdites structu-
    res dans l'image de projection.

2. Procédé pour visualiser une image en projection
2D selon la revendication 1, dans lequel le jeu de
données 3D est acquis par des techniques d'ima-
gerie par résonance magnétique à contraste de
phase.

3. Procédé pour visualiser une image en projection
2D selon la revendication 1, dans lequel le jeu de
données 3D est acquis par des techniques d'ima-
gerie par résonance magnétique à temps de vol.

4. Procédé pour visualiser une image en projection
2D selon la revendication 1, comprenant en outre
l'étape qui consiste à faire varier l'ordre statisti-
que pour un groupe prédéterminé de données tri-
dimensionnelles afin de produire des images
avec différents contrastes desdites structures
dans ledit objet dont on forme l'image.

5. Procédé pour visualiser une image en projection
2D selon la revendication 1, dans lequel le jeu de
données 3D est acquis par des techniques d'ima-
gerie de résonance magnétique utilisées pour
produire des images de tissus sensiblement im-
mobiles.

6. Procédé pour visualiser une image en projection
2D selon la revendication 1, dans lequel le jeu de
données 3D est acquis par des techniques d'ima-
gerie par tomographie.

7. Procédé pour visualiser une image en projection
2D selon la revendication 1, dans lequel le jeu de
données 3D est acquis par des techniques d'ima-
gerie par résonance magnétique.

Fig. 1

Fig. 2

Fig. 3

VOXEL VALUE

Fig. 4

Fig. 5

START

18 — Acquisition of Raw 3D data

19 — Create distribution histogram

21 — Determine and average all voxels intersecting projection ray

22 — calculate standard deviation for voxels intersecting projection

23 — raise standard deviation to a power (being the order of the statistic-1) to get normalization constant

24 — Subtract next voxel value from voxel average to get distribution difference

25 — Raise distribution difference to a power (being order of the statistic), to get weighted voxel difference

26 — Last voxel in projection ray?

27 — no

28 — Next voxel

yes 30

32 — Sum all weighted voxel differences in projection to get projection sum

33 — divide projection sum by the number of voxels in the projection then divide by the normalization constant to get a weighted projection value for one projection ray

34 — Last projection ray? — yes

35 — no

36 — Next projection ray

37

38

39 — Display all projections for a 2D image

Fig. 6